



# Calibrador A *plus*

Kit lot: 1803247180  
Calibrador A plus lot: 245460  
Reagent A lot: 247590

Clinical chemistry calibrator for autoanalyzers

## USES

The calibrator has been designed to be used with Wiener lab.'s reagents in autoanalyzers.

Check the value chart for the constituents, since they are lot specific.

## PRINCIPLE

**Calibrador A plus** contains the compounds usually determined in clinical chemistry laboratories.

It must be noted that the assigned values for the calibrator's different constituents have been obtained with Wiener lab.'s methods and reagents, thus, the obtained results will only be comparable to the list if the corresponding methods and reagents have been used.

## PROVIDED REAGENTS

**Calibrator:** lyophilized serum containing metabolites in the appropriate concentrations to ensure a correct calibration of autoanalyzers.

**A. Reagent A:** 25 mmol/l sodium carbonate solution, pH 10.

## INSTRUCTIONS FOR USE

- Open the vial, slowly pulling out the rubber stopper to avoid any loss of the lyophilized material.
- Add exactly 3 ml Reagent A (using a burette or a pipette with double calibration mark).
- Cap and mix by gentle inversion, avoiding the formation of foam. Do not shake.
- Let dissolve for 30 minutes at room temperature, mixing by inversion from time to time.
- Mix by inversion immediately before use.

## WARNINGS

Reagents are for "in vitro" diagnostic use.

This calibrator has been prepared from non-reactive material to HBsAg, HCV and HIV. However, it should be handled as if it were an infectious material.

Use the reagents according to the working procedures for clinical laboratories.

The reagents and samples should be discarded according to the local regulations in force.

## STABILITY AND STORAGE INSTRUCTIONS

The **Provided Reagents** are stable in refrigerator (2-10°C) until the expiration date shown on the box.

**Reconstituted Calibrator:** is stable for 8 hours at room temperature (<25°C), 2 days refrigerated (2-10°C) or 30 days frozen (-20°C).

In the dark, bilirubin is stable 4 hours at room temperature (<25°C), 8 hours refrigerated (2-10°C) or 2 weeks frozen (-20°C).

Avoid repeated freezing and thawing.

## INSTABILITY OR DETERIORATION OF REAGENTS

Any variation of the organoleptic features of the Calibrator may be a sign of its deterioration.

## PROCEDURE

Reconstituted Calibrator should be used in the same manner as an unknown sample, following the autoanalyzer parameters that come with the reagent kit in use.

## REFERENCE VALUES

Calibrator values have been established using Reference Material from the National Bureau of Standards, when it was possible.

## PROCEDURE LIMITATIONS

Mistakes in the reconstitution or conservation may cause erroneous results.

See PROCEDURE LIMITATIONS on the package insert of the kit in use.

## WIENER LAB. PROVIDES

- 2 x 3 ml (Cat. 1918005).

## REFERENCES

- Tietz, N. W. (Ed): Fundamentals of Clinical Chemistry, W. B. Saunders Co., Philadelphia, 865 (1982)
- Young, D.S., Effects of preanalytical variables on clinical laboratory test. AACC Press. Third ed. (2007).

**CALIBRADOR A PLUS METABOLITES ASSIGNED VALUES (37°C)**

<b>Analyte</b>	<b>Method</b>	<b>Value</b>	<b>Unit</b>
Uric acid	<b>Uricostat enzimático AA</b>	4,58	mg/dL
	<b>Uricostat enzimático AA líquida</b>	4,56	mg/dL
Albumin	<b>Albúmina AA</b>	3,58	g/dL
Direct bilirubin	<b>Bilirrubina</b>	2,54	mg/dL
	<b>Bilirrubina Directa AA</b>	2,49	mg/dL
	<b>Bilirrubina Directa AA líquida</b>	2,14	mg/dL
Total bilirubin	<b>Bilirrubina</b>	5,13	mg/dL
	<b>Bilirrubina Total AA</b>	4,20	mg/dL
	<b>Bilirrubina Total AA líquida</b>	4,25	mg/dL
Calcium	<b>Ca-Color AA</b>	10,6	mg/dL
	<b>Ca-Color Arsenazo III AA</b>	10,6	mg/dL
Cholesterol	<b>Colestat enzimático AA</b>	155	mg/dL
	<b>Colestat enzimático AA líquida</b>	155	mg/dL
Creatinine	<b>Creatinina cinética AA líquida - Conventional technique</b>	3,61	mg/dL
	<b>Creatinina cinética AA líquida - Compensate technique</b>	3,99	mg/dL
	<b>Creatinina enzimática AA líquida</b>	3,74	mg/dL
Phosphorus	<b>Fosfatemia UV AA</b>	5,92	mg/dL
Glucose	<b>Glicemia enzimática AA</b>	187	mg/dL
	<b>Glicemia enzimática AA líquida</b>	187	mg/dL
	<b>Glucose HK</b>	188	mg/dL
HDL Cholesterol	<b>HDL Colesterol Reactivo Precipitante</b>	171	mg/dL
	<b>HDL Colesterol FT</b>	545	mg/dL
Iron	<b>Fer-color AA</b>	203	µg/dL
	<b>Fer-color AA líquida</b>	186	µg/dL
Lactate	<b>Lactate</b>	30,4	mg/dL
LDL Cholesterol	<b>LDL Colesterol Reactivo Precipitante</b>	233	mg/dL
Lithium	<b>Lithium</b>	1,90	mmol/L
Magnesium	<b>Mg-Color AA</b>	2,90	mg/dL
	<b>Magnesium CPZ</b>	2,90	mg/dL
Total proteins	<b>Proteínas Totales AA</b>	5,18	g/dL
	<b>Total Protein</b>	5,10	g/dL
Triglycerides	<b>TG Color GPO/PAP AA</b>	128	mg/dL
	<b>TG Color GPO/PAP AA líquida</b>	128	mg/dL
UIBC	<b>UIBC/TIBC AA líquida</b>	158	µg/dL
Urea	<b>Urea UV cinética AA</b>	102	mg/dL
	<b>Urea UV cinética AA líquida</b>	102	mg/dL

**CALIBRADOR A PLUS ENZYMES ASSIGNED VALUES (37°C)**


<b>Analyte</b>	<b>Method</b>	<b>Value</b>	<b>Unit</b>
Alanine aminotransferase (GPT/ALT)	<b>GPT (ALT) UV AA</b>	111	U/L
	<b>GPT (ALT) UV AA líquida</b>	111	U/L
Amylase	<b>Amilasa 405 cinética AA</b>	258	U/L
	<b>Amilasa 405 cinética AA líquida</b>	258	U/L
Aspartato aminotransferase (GOT/AST)	<b>GOT (AST) UV AA</b>	112	U/L
	<b>GOT (AST) UV AA líquida</b>	112	U/L
Cholinesterase	<b>Colinesterasa AA</b>	5100	U/L
	<b>Cholinesterase</b>	4980	U/L
Creatine kinase	<b>CK-NAC UV AA</b>	308	U/L
	<b>CK-NAC UV AA líquida</b>	308	U/L
Creatine kinase - MB	<b>CK-MB NAC UV AA líquida</b>	378	U/L
Acid phosphatase	<b>Fosfatasa Acida Total y Prostática cinética - FANP</b>	9,22	U/L
	<b>Fosfatasa Acida Total y Prostática cinética - FAT</b>	25,0	U/L
Alkaline phosphatase	<b>ALP 405 AA líquida</b>	440	U/L
$\gamma$ -Glutamyl transferase	<b><math>\gamma</math>-G test cinética AA</b>	120	U/L
	<b><math>\gamma</math>-G test cinética AA líquida</b>	120	U/L
Lactate deshydrogenase	<b>LDH-P UV AA</b>	515	U/L
	<b>LDH-P UV AA líquida</b>	515	U/L
	<b>LDH-L</b>	263	U/L
Lipase	<b>Lipasa AA líquida</b>	101	U/L

# Symbols

The following symbols are used in the packaging for Wiener lab. diagnostic reagent kits.



This product fulfills the requirements of the European Directive 98/79 EC for "in vitro" diagnostic medical devices



Authorized representative in the European Community



"In vitro" diagnostic medical device



Contains sufficient for <n> tests



Use by



Temperature limitation (store at)



Do not freeze



Biological risks



Volume after reconstitution



Contents



Batch code



Manufactured by:



Harmful



Corrosive / Caustic



Irritant



Consult instructions for use



Calibrator



Control




Positive Control



Negative Control



Catalog number

 Wiener Laboratorios S.A.I.C.  
Riobamba 2944  
2000 - Rosario - Argentina  
<http://www.wiener-lab.com.ar>  
Tech.Dir.: Viviana E. Cétola  
Biochemist  
A.N.M.A.T. Registered product  
Cert. N°: 1976/97



**Wiener lab.**

2000 Rosario - Argentina