

CK-MB

Control - 3 niveles

LOT 2001359320

LOT Nivel 1 354130

LOT Nivel 2 354140

LOT Nivel 3 354150

VALORES ASIGNADOS

VALORES ASIGNADOS DE CK-MB (U/l)* - BT 3000 PLUS / WIENER LAB, CB SERIES									
KIT	NIVEL 1			NIVEL 2			NIVEL 3		
	VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE	
CK-MB DS	30.7	21.5	39.9	43.7	32.8	54.6	88.4	66.3	110.5
CK-MB LÍQ.	31.3	21.9	40.7	46.3	34.7	57.9	92.1	69.1	115.1

VALORES ASIGNADOS DE CK-MB (U/l)* - KONELAB SERIES / WIENER LAB. CT SERIES									
KIT	NIVEL 1			NIVEL 2			NIVEL 3		
	VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE	
CK-MB DS	35.3	24.7	45.9	51.0	38.3	63.8	105	79	131
CK-MB LÍQ.	34.9	24.4	45.4	48.2	36.2	60.3	99.5	74.6	124.4

VALORES ASIGNADOS DE CK-MB (U/l)* - METROLAB 2300 PLUS / WIENER LAB. CM SERIES									
KIT	NIVEL 1			NIVEL 2			NIVEL 3		
	VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE	
CK-MB DS	28.5	20.0	37.1	41.9	31.4	52.4	83.6	62.7	104.5
CK-MB LÍQ.	36.8	25.8	47.8	51.1	38.3	63.9	96.3	72.2	120.4

VALORES ASIGNADOS DE CK-MB (U/l)* - TARGA BT 3000									
KIT	NIVEL 1			NIVEL 2			NIVEL 3		
	VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE	
CK-MB DS	30.7	21.5	39.9	43.7	32.8	54.6	88.4	66.3	110.5
CK-MB LÍQ.	31.3	21.9	40.7	46.3	34.7	57.9	92.1	69.1	115.1

VALORES ASIGNADOS DE CK-MB (U/l)* - WIENER LAB. CMD 800 SERIES									
KIT	NIVEL 1			NIVEL 2			NIVEL 3		
	VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE		VALOR MEDIO	RANGO ACCEPTABLE	
CK-MB LÍQ.	28.5	20.0	37.1	43.7	32.8	54.6	91.7	68.8	114.6

* a 37°C

** Valores obtenidos empleando técnica con calibración / Valores obtidos utilizando técnica com calibração / Values obtained using calibration technique / Wartości uzyskane stosując techniki z kalibracją

Glosario/Glossário/Glossary/Glosariusz

Español	Português	English	Polski
kit	kit	kit	zestaw
nivel 2	nível 2	level 2	poziom 2
nivel 3	nível 3	level 3	poziom 3
valores asignados	valores estabelecidos	assigned values	zakres wartości
valor medio	valor médio	mean	średnia wartość
rango acceptable	faixa aceitável	acceptable range	pożądany zakres

