

Control Plasma

Pathologic

LOT

2007383090

Reference Values

Reagent	Manual Technique		Automated Technique*	
Prothrombin Time (sec)	24.8	41.4	24.8	41.4
APTT (sec)	47.8	79.6	-	-
APTT ellagic (sec)	41.3	68.8	40.7	67.9
Fibrinogen (mg/dl)	120	180	120	180
Thrombin Time (sec)				
4.0 UNIH/ml	12.3	18.5	12.3	18.5
2.7 UNIH/ml	15.4	23.0	15.0	22.4

(*) Coa DATA coagulation instruments

Dutch Diagnostics BV

Hekkehorst 3 - 7207 BS Zutphen

The Netherlands

Phone: +31 575 512 300

Fax: +31 575 510 164

E-mail: info@dutch-diagnostics.com