

IgE Control

Turbitest AA

LOT 2108445910

LOT Nivel 1 445870

LOT Nivel 2 445880

VALORES ASIGNADOS

BT 3000 PLUS / WIENER LAB. CB SERIES						
KIT	NIVEL 1			NIVEL 2		
	VALOR MEDIO (UI/ml)	RANGO ACEPTABLE		VALOR MEDIO (UI/ml)	RANGO ACEPTABLE	
IgE Control	64,8	45,4	84,2	420	315	525

KONELAB SERIES / WIENER LAB. CT SERIES						
KIT	NIVEL 1			NIVEL 2		
	VALOR MEDIO (UI/ml)	RANGO ACEPTABLE		VALOR MEDIO (UI/ml)	RANGO ACEPTABLE	
IgE Control	63,7	47,8	79,6	404	303	505

METROLAB 2300 PLUS / WIENER LAB. CM SERIES						
KIT	NIVEL 1			NIVEL 2		
	VALOR MEDIO (UI/ml)	RANGO ACEPTABLE		VALOR MEDIO (UI/ml)	RANGO ACEPTABLE	
IgE Control	59,8	44,9	74,8	379	284	474

WIENER LAB. CMD SERIES						
KIT	NIVEL 1			NIVEL 2		
	VALOR MEDIO (UI/ml)	RANGO ACEPTABLE		VALOR MEDIO (UI/ml)	RANGO ACEPTABLE	
IgE Control	66,4	49,8	83,0	409	307	511

Glosario/Glossário/Glossary/Glosariusz

Español	Português	English	Polski
kit	kit	kit	zestaw
nivel 1	nível 1	level 1	poziom 1
nivel 2	nível 2	level 2	poziom 2
valores asignad	valores estabelecidos	assigned values	zakres wartości
valor medio	valor médio	mean	średnia wartość
rango acceptable	faixa aceitável	acceptable range	pożądany zakres

