

Proti U/LCR

Control - 2 niveles

LOT 2001356580

LOT Nivel 1 352110

LOT Nivel 2 352120

VALORES ASIGNADOS PARA TECNICA MANUAL Y ANALIZADORES						
TECNICA/ ANALIZADOR	NIVEL 1			NIVEL 2		
	VALOR MEDIO (mg/dL)	RANGO (mg/dL)		VALOR MEDIO (mg/dL)	RANGO (mg/dL)	
Konelab Series / Wiener lab, CT Series	30.3	22.7	37.8	102	81	122
Metrolab 2300 plus / Wiener lab. CM Series	32.7	24.5	40.9	100	80	120
Targa BT 3000	33.1	24.8	41.3	98.5	73.9	123.1
BT 3000 plus / Wiener lab. CB Series	33.1	24.8	41.3	98.5	73.9	123.1
Wiener lab. CMD Series	31.3	23.4	39.1	101	80	121
Técnica Manual	32.8	24.6	41.0	98	74	123

Glosario/Glossário/Glossary/Glosariusz

Español	Português	English	Polski
nivel	nível	level	poziom
fecha	data	date	data
valores asignados	valores estabelecidos	assigned values	zakres wartości
valor medio	valor médio	mean	średnia wartość
rango	faixa	range	zakres
técnica manual	técnica manual	manual technique	metody manualne
analizadores	analísadores	analyzers	analizator

